

## Calibration Certificate

Date: 6/3/2011

Certificate #: 831104

### Calibration Performed By:

MSI-VIKING GAGE, LLC  
321 TUCAPAU ROAD  
DUNCAN, S.C. 29334

### For:

DIGIPAS USA  
304 W. MAIN STREET #120  
AVON, CT 06001

### Equipment Information

I.D.: DWL-1000XY  
Description: DIGITAL LEVEL  
Manufacturer: DIGIPAS  
Gage Type: DIGITAL LEVEL  
Temp./RH: 68 F / 40  
Cal. Interval: 1 YEARS  
Cal Date: 6/3/2011  
Customer PO: CC

Serial Number: DWL-1000XY  
Model Number: ~~DWL-600E~~  
Unit of Measure: DEGREE  
Performed By: JSINARD  
Cal. Due Date: 6/3/2012  
Received Condition: GOOD  
Resulting Condition: IN TOLERANCE  
Department:

### Calibration Notes

UNCERTAINTY +/- .01 DEGREE.

### Test Points

Seq.	Description	Standard	Tolerance -	Tolerance +	As Found	Final	Attribute 1	Attribute 2
1	POSITION 1	0.000000	-0.100000	0.100000	0.000000	0.000000		
2	POSITION 2	0.000000	-0.100000	0.100000	0.000000	0.000000		
3	POSITION 3	0.000000	-0.100000	0.100000	0.000000	0.000000		
4	POSITION 4	0.000000	-0.100000	0.100000	0.000000	0.000000		
5	POSITION 5	0.000000	-0.100000	0.100000	0.000000	0.000000		
6	POSITION 6	0.000000	-0.100000	0.100000	0.000000	0.000000		

### Standards Used To Calibrate Equipment

Company	I.D.	Description	Last Cal.	Due Date	NIST #
MSI-VIKING GAGE, LLC	010320	GAGE BLOCK SET 81 PC. SQUARE	4/15/2011	4/15/2012	821/279484-10
MSI-VIKING GAGE, LLC	V020800-1	SURFACE PLATE 18X24	10/15/2010	10/15/2011	821/263310-00
MSI-VIKING GAGE, LLC	VKG-085	SINE BAR	9/13/2010	9/13/2012	821/271641-05

## Calibration Certificate

Date: 6/3/2011

Certificate #: 831104

### Procedures Used In This Event:

Procedure Company	Procedure Name	Description	Revision Level	Revision Date
MSI-VIKING GAGE, LLC	VCP-M-1-C012	LEVELS	3	11/23/2004

The MSI-Viking Gage, LLC. Calibration Laboratory operates under Standard Operating Procedures intended to implement the requirements of ISO 10012 and ISO/IEC 17025, and has been accredited to ISO/IEC 17025 by A2LA, Certificate #1387.01. All calibrations are traceable to the National Institute of Standards and Technology (NIST). The reported uncertainty is based on a standard of uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95%. This certificate of calibration is neither a contract nor a contractual representation. MSI-Viking Gage is responsible for the calibration process results during the time of the calibration process. Any number of factors may cause the calibration item to drift out of calibration before the calibration interval has expired. It shall be the customer's responsibility to ensure that the TI's remain within calibration for the duration of the assigned calibration interval. This certificate of calibration may not be copied except in full and only with the written consent of MSI-Viking Gage, LLC. See "Scope of Accreditation" for accredited services.